


## VERIFICATION OF COMPLIANCE

**Issue Date:** April 11, 2023  
**Applicant:** ELJINTEK, INC.  
**Address:** 6F-1 No.286-3 Shin-Ya Road Chien-Chen District Kaohsiung Taiwan 806  
**Manufacturer:**  
(1) ELJINTEK, INC.  
(2) Youngyear Electronics Co., Ltd.  
(3) Maxim Enterprise Co Ltd. J&C Electric Mfg. Co.  
(4) Mean Well Technology Co., Ltd.  
(5) SuZhou Mean Well Technology Co., Ltd.  
(6) Mitra Power Solution Co., Ltd.  
(7) Dongguan Shengfone Electronics Technology Co.,Ltd  
**Address:**  
(1) 6F-1 No.286-3 Shin-Ya Road Chien-Chen District Kaohsiung Taiwan 806  
(2) Jin Zhu Industries Area Qing Xi Town Dongguan , Guangdong 523646 China  
(3) No. 134-1 Puxinhu Rd. Tangxia Zhen Dong Guan Shi, Guang Dong, China  
(4) No. 28, Wuquan 3rd Rd.,Wugu Dist.,New Taipei City 24891 Taiwan, R.O.C.  
province China.  
(5) No. 77, Jian-min Road, Dong-qiao, Pan-yang Ind. Park, Huang-dai Town, Xiang-cheng Dist., Suzhou, Jiangsu 215152, China.  
(6) First Industrial district,Cunwei Village,HengLi Town DongGuan GuangDong PRC.  
(7) Xinlide Industrial Estate, Dongfeng Road, Qingxi Town, Dongguan City, Guangdong, P.R. China  
**Product:** SWITCHING POWER SUPPLY ; AC/DC SWITCHING ADAPTOR ; MEDICAL ADPATOR  
**Brand Name/Trade Mark:** ELJINTEK, INC.  
**Model/Type:** GMPU06x-yzwzzzzz; GSM06xbzwzzzzz; GEM06lbzwzzzzz (x= U, E, SA, BS, AR, KR, CH, JP ,UI , EI, SAI, BSI, ARI, KRI, CHI, JPI; y= 1, 2, 3, 4, 5, 6, 1-1 or 11; b= 05, 06, 07, 09, 12, 15, 18 or 24; z= 0 to 9, A to Z; hyphen or blank; w= USB)  
**Added Model(s):** N/A  
**Applicable Standards:**  
**EN 60601-1-2: 2015 + A1: 2021**  
(Including subclause 8.11 immunity to proximity magnetic fields, Test method refer to IEC 61000-4-39)  
**EN 55011: 2016 +A1: 2017+A11: 2020**  
EN 61000-3-2: 2014, EN 61000-3-3: 2013  
IEC 61000-4-2: 2008, IEC 61000-4-3: 2020  
IEC 61000-4-4: 2012, IEC 61000-4-5: 2014 + A1: 2017  
IEC 61000-4-6: 2013 + COR1: 2015, IEC 61000-4-8: 2009  
IEC 61000-4-11: 2020 + COR1: 2020 + COR2: 2022  
**Test Laboratory:** Compliance Certification Services Inc.  
Tainan Lab  
No. 168, Ln. 523, Sec. 3, Zhongzheng Rd., Rende Dist., Tainan City, 717017, Taiwan  
**Test Report No. :** TMTN2302000207NE dated on April 11, 2023

**Conclusion:**

Based upon a review of the Test Report(s), the tested sample of the product mentioned above is deemed to comply with the requirements of the above standards.

**Note:** This verification is only valid for the product and configuration described and in conjunction with the test report as detailed above.

**Authorised Signatory:****Compliance Certification Services Inc.****Jeter Wu**  
**Manager****FM-023A-R05**

This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com.tw/Terms-and-Conditions> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com.tw/Terms-and-Conditions>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of client's instruction, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced, except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.