


VERIFICATION OF COMPLIANCE

Issue Date: April 27, 2023
Applicant: ELJINTEK ,INC.
Address: 6F-1, No.286-3, Shin Ya Road, Chien Chen Dist.,Kaohsiung 806 Taiwan
Manufacturer:
(1) ELJINTEK,INC.
(2) Mean Well Enterprises Co, Ltd
(3) Mitra Power Solution Co.,Ltd
(4) YOUNGYEAR ELECTRONICS CO LTD
Address:
(1) 6F-1, No.286-3, Shin Ya Road, Chien Chen Dist., Kaohsiung 806 Taiwan
(2) No. 28, Wuquan 3rd Rd.,Wugu Dist.,New Taipei City 248 Taiwan
(3) First Industrial district,Cunwei Village,HengLi Town DongGuan GuangDong PRC.
(4) Jin Zhu Industries Area Qing Xi Town Dongguan , Guangdong 523646 China
Product:
Brand Name/Trade Mark: ELJINTEK ,INC.
Model/Type: GMPU12x-y & GMPU12x-y-USB;
(X= U, E, UI, EI, SAI, BSI, ARI, KRI, CHI ,JPI; Y= 1 to 8 or 1-1); "x" for
different plug type; "U" or "UI" denotes for American type or Interchangeable American type;; "E"
or "EI" denotes for Europe type or Interchangeable Europe type;; "SAI" denotes for 2 Pin
Interchangeable Australia type;; "BSI" denotes for Interchangeable UK type.; "ARI" denotes for
Interchangeable Argentina type;; "KRI" denotes for Interchangeable Korea type;; "CHI" denotes for
Interchangeable china type;; "JPI" denotes for
Interchangeable Japan type;; "Y" for output rating
GMPU12x-y & GMPU12x-y-USB
(x= U, E, UI, EI, SAI, BSI, ARI, KRI, CHI,JPI; y= 1 to 8 or 1-1)
GSM12xyzzzz & GSM12xyzzzz-USB
("I" denotes for Interchangeable type ; y= 05,07,09,12,15,18,24,28 or 48; z= 0 to 9, A to Z,
hyphen or blank)
GEM12lyzzzz & GEM12lyzzzz-USB ("I" denotes for Interchangeable type;
y=05.07.09.12.15.18.24.28 or 48; z= 0 to 9, A to Z, hyphen or blank)
Applicable Standards:
EN 60601-1-2: 2015 + A1: 2021
(Including subclause 8.11 immunity to proximity magnetic fields, Test method refer to IEC 61000-
4-39.)
EN 55011: 2016+A1: 2017+A11: 2020
EN 61000-3-2: 2014, EN 61000-3-3: 2013
IEC 61000-4-2: 2008, IEC 61000-4-3: 2020
IEC 61000-4-4: 2012, IEC 61000-4-5: 2014+A1: 2017
IEC 61000-4-6: 2013+C1: 2015, IEC 61000-4-8: 2009
IEC 61000-4-11: 2020+C1: 2020+C2: 2022
Test Laboratory:
Compliance Certification Services Inc.
Tainan Lab
No. 168, Ln. 523, Sec. 3, Zhongzheng Rd., Rende Dist., Tainan City, 717017, Taiwan
Test Report No. : TMTN2302000216NE dated on April 27, 2023
Conclusion:
Based upon a review of the Test Report(s), the tested sample of the product
mentioned above is deemed to comply with the requirements of the above standards.
Note: This verification is only valid for the product and configuration described and in conjunction
with the test report as detailed above.

Authorised Signatory:

Compliance Certification Services Inc.

Jeter Wu
Manager

FM-023A-R05

This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com.tw/Terms-and-Conditions> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com.tw/Terms-and-Conditions>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of client's instruction, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced, except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.