

Eljintek, Inc.
Ms. Lydia Zhang

6F-1, No.286-3, Shin Ya Road
Chien Chen Dist.
Kaohsiung 806
Taiwan

Date : 22.04.2022
Our ref. : ZTW1
Your ref.: P00630773

Ref : TA License

Type of Equipment : SWITCHING POWER SUPPLY
Model Designation : See Certificate
Certificate No. : TA 50373951 0002
Report No. : 50071022 006

Dear Ms. Lydia Zhang,

The above specified technical equipment has been tested and found to be in accordance with the relevant requirements.

Enclosed you will find the certificate of approval.

If cancellation of the certificate is submitted by November 15 in a given year, no fee will be charged for the following year.

We appreciate your kind support and would like to offer our assistance and continuous services in the future.

Please contact our office for approval of your new products.

With kind regards,

Certification Body


Dipl.-Ing. W. Hsu

Enclosure

TÜV RHEINLAND TAIWAN LTD.

11F., No. 758, Sec. 4, Bade Rd.,
Songshan Dist., Taipei City 105,
Taiwan, R. O. C.
Tel. (02) 2172-7000
Fax (02) 2172-1322
<http://www.tuv.com>

NEW TAIPEI BRANCH:
No. 458-18 & 458-19,
Sec. 2, Fenliao Rd., Linkou Dist.,
New Taipei City 244,
Taiwan, R. O. C.
Tel. (02) 2172-1000
Fax (02) 2172-1322

TAICHUNG BRANCH:
3F., No. 497, Zhongming South Rd.,
West Dist., Taichung City 403,
Taiwan, R. O. C.
Tel. (02) 2172-7000
Fax (02) 2172-1322

KAOHSIUNG BRANCH:
27F.-3, No.80, Minzu 1st Rd.,
Sanmin Dist., Kaohsiung City 807,
Taiwan, R. O. C.
Tel. (07) 262-6000
Fax (07) 380-0577

Certificate



Certificate no.

TA 50373951 01

License Holder:

Eljintek, Inc.
6F-1, No.286-3, Shin Ya Road
Chien Chen Dist.
Kaohsiung 806
Taiwan

Manufacturing Plant:

Eljintek, Inc.
6F-1, No.286-3, Shin Ya Road
Chien Chen Dist.
Kaohsiung 806
Taiwan

Test report no.: ZTW1- 50071022 003

Client Reference: EH/SPC-1606022-app

Tested to: IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1+A12
IEC 60601-1-11:2015
EN 60601-1-11:2015

Certified Product: (SWITCHING POWER SUPPLY)

License Fee - Units

Type Designation	: GMPU12x-y	8
x stands for	: E or EI	1
y stands for	: 1, 1-1, 2, 3, 4, 5, 6, 7, 8	1
Rated Voltage	: AC 100-240V, 50/60Hz	
Rated Current	: 0.4-0.2A	
Output Voltage	: see constructional dataform	
Output Current	: see constructional dataform	
max. Ambient Temperature	: 40°C	
Protection Class	: II	

This certificate is valid for maximum 5 years.

Appendix: 1

10

Licensed Test mark:



TÜV Rheinland Taiwan Ltd

Signature

Dipl.-Ing. W. Hsu



Date of Issue
(day/mo/yr)

15/03/2017

Certificate



Certificate no.

TA 50373951 02

License Holder:

Eljintek, Inc.
6F-1, No.286-3, Shin Ya Road
Chien Chen Dist.
Kaohsiung 806
Taiwan

Manufacturing Plant:

Eljintek, Inc.
6F-1, No.286-3, Shin Ya Road
Chien Chen Dist.
Kaohsiung 806
Taiwan

Test report no.: ZTW1- 50071022 006

Client Reference: P00630773

Tested to: IEC 60601-1:2005+A1+A2
EN 60601-1:2006+A11+A1+A12+A2
IEC 60601-1-11:2015+A1
EN 60601-1-11:2015+A1

Certified Product: (SWITCHING POWER SUPPLY)

License Fee - Units

as page 01

Change

Test Requirement : see above

Remark: 1. The risk management requirements and usability of the standard were not addressed.
2. This certificate is valid for maximum 5 years.

Appendix: 2

Licensed Test mark:



IEC 60601-1
EN 60601-1
IEC 60601-1-11
EN 60601-1-11

www.tuv.com
ID 1419045372

TÜV Rheinland Taiwan Ltd

Signature

Dipl.-Ing. W. Hsu



Date of Issue
(day/mo/yr)

22/04/2022