

## VERIFICATION OF COMPLIANCE

<b>Issue Date:</b>	April 10, 2023
<b>Applicant:</b>	ELJINTEK, INC.
<b>Address:</b>	6F-1 No.286-3 Shin-Ya Road Chien-Chen District Kaohsiung Taiwan 806
<b>Manufacturer:</b>	1) ELJINTEK INC 2) Mean Well Enterprises Co, Ltd 3) Mitra Power Solution Co., Ltd 4) YOUNGYEAR ELECTRONICS CO LTD 5) Dongguan Shengfone Electronics Technology Co., Ltd.
<b>Address:</b>	1) 6F-1, No.286-3, Shin Ya Road, Chien Chen Dist., Kaohsiung 806 Taiwan 2) No. 28, Wuquan 3rd Rd., Wugu Dist., New Taipei City 248 Taiwan 3) First Industrial district, Cunwei Village, HengLi Town DongGuan GuangDong PRC. 4) Jin Zhu Industries Area Qing Xi Town Dongguan, Guangdong 523646 China 5) Shengfeng Electronic Factory, Dongfeng RD., Qingxi Town, Dongguan city, Guangdong Province, China
<b>Product:</b>	Switching Power Supply
<b>Brand Name/Trade Mark:</b>	<b>ELJINTEK, INC.</b>
<b>Model/Type:</b>	GMPU60x-yzzzzz series (x = B (denote IEC 60320 C8 inlet type) or U, E, SA, BS, AR, KR, CH, UI, EI, SAI, BSI, ARI, KRI, CHI (denotes AC plug type); y = 1, 2, 3, 4, 5, 6, 7, 8, 1-1 or 3A; z = 0 to 9, A to Z, hyphen or blank)
<b>Added Model(s):</b>	GSM60xbzzzzz (x = B (denotes AC plug type) or U, E, AU, UK, AR, KR, CH; b = 05, 07, 09, 12, 15, 18, 24, 28, 48 or 12A; z = 0 to 9, A to Z, hyphen or blank); GEM60lbzzzzz (b = 05, 07, 09, 12, 15, 18, 24, 28, 48 or 12A; z = 0 to 9, A to Z, hyphen or blank)
<b>Applicable Standards:</b>	IEC 60601-1-2: 2014 + A1: 2020 EN 60601-1-2: 2015 + A1: 2021 CISPR 11: 2015 + A1: 2016 + A2: 2019 (Ed 6.2) IEC 61000-3-2: 2018 IEC 61000-3-3: 2013 + A1: 2017 IEC 61000-4-2: 2008 IEC 61000-4-3: 2006 + A1: 2007 + A2: 2010 IEC 61000-4-4: 2012 IEC 61000-4-5: 2014 + A1: 2017 IEC 61000-4-6: 2013 + COR1: 2015 IEC 61000-4-8: 2009 IEC 61000-4-11: 2004 + A1: 2017 IEC 60601-1-2: 2014 + A1: 2020 Subclause 8.11 (IEC 61000-4-39 for §8.10)
<b>Test Laboratory:</b>	Compliance Certification Services Inc. Xindian Lab No. 163-1, Jhongsheng Rd., Xindian Dist., New Taipei City, Taiwan
<b>Test Report No.:</b>	TMXD2302000517DE, dated on April 10, 2023
<b>Conclusion:</b>	Based upon a review of the Test Report(s), the tested sample of the product mentioned above is deemed to comply with the requirements of the above standards.
<b>Note:</b>	This verification is only valid for the product and configuration described and in conjunction with the test report as detailed above.

### Authorised Signatory:



Compliance Certification Services Inc.

Jason Lee

Section Manager

This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com.tw/Terms-and-Conditions> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com.tw/Terms-and-Conditions>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of client's instruction, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced, except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.